
START
11
WAS AED ACTIVATED BY BUTTON?
YES
NO
13a
ISSUE INITIAL PROMPT
13b
PAUSE TO ALLOW REMOVAL OF CLOTHING
13c
ISSUE SECOND PROMPT
ARE PADS BEING HANDLED?
13d
YES
NO
HAVE THESE PROMPTS BEEN GIVEN MORE THAN "X" TIMES
13e
YES
NO
12a
ISSUE INITIAL PROMPT
12b
PAUSE TO ALLOW REMOVAL OF CLOTHING
ISSUE SECOND PROMPT
12c
12d
ARE PADS BEING HANDLED?
NO
YES
ISSUE PICTURE PROMPT
14
HAS PAD BEEN REMOVED FROM LINER?
15
YES
NO
16a
ISSUE PAD APPLICATION PROMPT
17
ARE BOTH PADS PLACED?
YES
NO
ISSUE PAD CORRECTION PROMPT
16b
ANALYZE ECG SIGNAL
18
IS ECG SIGNAL ERRATIC?
19
YES
NO
20
ISSUE PAD CORRECTION PROMPT
END
10

FIG. 1